

**Blind separation of analytes in nuclear magnetic resonance spectroscopy and mass spectrometry: sparseness-based robust multicomponent analysis**

Journal:	<i>Analytical Chemistry</i>
Manuscript ID:	ac-2009-02640y.R1
Manuscript Type:	Article
Date Submitted by the Author:	
Complete List of Authors:	Kopriva, Ivica; Rudjer Boskovich Institute, Laser and Atomic R&D Jeric, Ivanka



Submitted to *Analytical Chemistry*

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

# Blind separation of analytes in nuclear magnetic resonance spectroscopy and mass spectrometry: sparseness-based robust multicomponent analysis<sup>§</sup>

*Ivica Kopriva<sup>a</sup> and Ivanka Jerić<sup>b</sup>*

Ruđer Bošković Institute, Bijenička cesta 54, HR-10000, Zagreb, Croatia

<sup>a</sup>Division of Laser and Atomic Research and Development

<sup>b</sup>Division of Organic Chemistry and Biochemistry

ikopriva@irb.hr

Tel.: +385-1-4571-286. Fax: +385-1-4680-104

<sup>§</sup>Patent pending under the number PCT/HR2009/00028.

1 **Metabolic profiling of biological samples involves nuclear magnetic resonance (NMR)**  
2 **spectroscopy and mass spectrometry coupled with powerful statistical tools for complex data**  
3 **analysis. Here, we report a robust, sparseness-based method for the blind separation of analytes**  
4 **from mixtures recorded in spectroscopic and spectrometric measurements. The advantage of the**  
5 **proposed method in comparison to alternative blind decomposition schemes is that it is capable of**  
6 **estimating the number of analytes, their concentrations, and the analytes themselves, from**  
7 **available mixtures only. The number of analytes can be less than, equal to or greater than the**  
8 **number of mixtures. The method is exemplified on blind extraction of four analytes from three**  
9 **mixtures in 2D NMR spectroscopy and five analytes from two mixtures in mass spectrometry. The**  
10 **proposed methodology is of widespread significance for natural products research and the field of**  
11 **metabolic studies, whereupon mixtures represent samples isolated from biological fluids or tissue**  
12 **extracts.**  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28

29 Current achievements and progress in the field of systems biology and functional genomics depend  
30 sensitively on the level of development of associated analytical techniques.<sup>1</sup> Metabolic profiling of  
31 biological fluids, cells and tissues provides insight into physiological processes, where it addresses  
32 multiple aims. These include disease diagnostics, xenobiotic toxicity, and nutrition- and environmental-  
33 influenced responses of living systems.<sup>2</sup> Information-rich techniques such as NMR spectroscopy and  
34 mass spectrometry (MS) represent powerful diagnostic tools for metabolomic and metabonomic studies,  
35 particularly through the identification and quantification of chemical entities directly correlated with  
36 certain disorder or disease (biomarkers).<sup>3</sup> One of the main disadvantages of <sup>1</sup>H NMR spectroscopy is  
37 signal overlapping, which increases with the number of components, their complexity and/or similarity.  
38 This shortcoming can be significantly reduced by spreading to the second dimension. While 2D NMR  
39 spectroscopy is commonly used for the structure elucidation of biomacromolecules, there are limited  
40 examples of its application in metabolic analysis. 2D homonuclear and heteronuclear NMR  
41 spectroscopy was applied to the studies of central nervous system and muscles,<sup>4</sup> but recently also to  
42 analyze healthy and cancerous tissues.<sup>5</sup> Despite significant improvement in many aspects, through  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

1 isotopic labeling and chemoselective tagging,<sup>6</sup> 2D NMR spectra are still challenged by limited  
2 resolution. Thus, a high level of data complexity generated in metabolic studies requires adequate data  
3 analysis. A multivariate data analysis methodology capable of blind extraction of a single component  
4 (analyte) spectrum out of a mixture would significantly improve and accelerate metabolic fingerprinting,  
5 biomarker searches and natural products analysis. Known as blind source separation (BSS), it has been  
6 reported previously in NMR,<sup>7</sup> infrared (IR)<sup>8</sup>, electronic paramagnetic resonance (EPR)<sup>9</sup> and Raman<sup>10</sup>  
7 spectroscopy as well as mass spectrometry.<sup>11</sup> However, in all these examples, algorithms of independent  
8 components analysis (ICA)<sup>12</sup> were used. These techniques assume that components are statistically  
9 independent and their number is less than or equal to the number of mixtures available. When the  
10 spectra of different analytes overlap significantly, the statistical independence assumption is only  
11 partially fulfilled,<sup>11</sup> causing ICA to fail. Moreover, ICA can not solve BSS problems characterized by  
12 more components than available mixtures. When mixtures represent samples of biological fluids, plant  
13 or tissue extracts with a few hundreds of analytes, overlapping of resonant peaks is a common  
14 phenomenon.<sup>3</sup> Moreover, the number of components present and their concentrations are not known in  
15 advance. This adversely affects the accuracy of the ICA-based blind extraction of analytes. The same  
16 comment applies to the band target entropy method (BTEM) applied to the analysis of multicomponent  
17 2D NMR spectra.<sup>13</sup> These lead to the underdetermined blind source separation (uBSS) scenario,<sup>12,14,15</sup>  
18 where an *unknown* number of components ought to be extracted, having only a *smaller* number of  
19 mixtures spectra at ones disposal, whereas the number of mixtures ought to be greater than one. We  
20 have recently demonstrated blind extraction of analytes from a smaller number of mixtures in FT-IR  
21 spectroscopy,<sup>16</sup> mass spectrometry<sup>17</sup> and <sup>1</sup>H and <sup>13</sup>C NMR spectroscopy,<sup>18</sup> exploiting sparseness  
22 between the components in some representation domain. However, the mutual sparseness assumption is  
23 severely violated when the number of components, their complexity or their similarity increases, leaving  
24 us unable to deal with biologically relevant problems. Here, we have proposed and verified an approach  
25 toward the solution of a problem considered within the chemometrics community "too ill-posed and  
26 thus, unsolvable."<sup>13</sup> The method relies on an assumption that components are mutually sparse (do not  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

1 overlap) at a *small number of points only*. Thus, it is well founded to expect it to be successful in blind  
2 extraction of analytes from a small number of complex mixtures. In combination with daily  
3 improvements of analytical tools, this approach could potentially yield a viable method for biomarker  
4 identification and extraction from biological samples. The method is exemplified on blind extraction of  
5 analytes from the mixtures recorded in 2D NMR spectroscopy and mass spectrometry. Since it is  
6 derived to solve general-purpose BSS problems, its applications clearly extend beyond the blind  
7 extraction of analytes. As but one application example of great importance in systems biology, we point  
8 to the reconstruction of transcription factors in gene regulating networks.<sup>3,19,20</sup>

## 21 THEORY AND ALGORITHM

22 **Linear mixture model.** Blind extraction of the analytes is based upon the linear mixture model (1)<sup>7-11</sup>,

23  
24  
25  
26 16-18

$$30 \quad \mathbf{X} = \mathbf{AS} \quad (1)$$

31  
32  
33  
34  
35 where  $\mathbf{X} \in C^{I_n \times I_1 I_2 \dots I_{n-1}}$  represents a matrix of, in the general case, complex data. The  $I_n$  rows contain  
36 mixtures measured by some  $(n-1)$ -dimensional spectroscopic modality;  $\mathbf{A} \in R_{0+}^{I_n \times J}$  is an unknown  
37 nonnegative real matrix of concentration profiles of the unknown number of  $J$  analytes; and  
38  $\mathbf{S} \in C^{J \times I_1 I_2 \dots I_{n-1}}$  is a matrix of (potentially complex) data, the  $J$  rows of which contain analytes. The linear  
39 mixtures model (1) is verified in ref. 17 to be a valid description of the mixtures of analytes mass  
40 spectra. The adopted notation can be used to represent mixtures measured either from one-dimensional  
41 or from multi-dimensional spectroscopic or spectrometric modalities, which might be necessary if the  
42 complexity of analytes is very high and/or their number is very large. According to the standard notation  
43 adopted for use in multiway analysis,<sup>21</sup> mixtures recorded by  $(n-1)$ -dimensional modalities actually form  
44 the  $n$ -dimensional tensor:  $\underline{\mathbf{X}} \in C^{I_1 \times I_2 \times \dots \times I_n}$ . The two-dimensional representation  $\mathbf{X}$  adopted by the model  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

(1) is obtained from a tensor  $\underline{\mathbf{X}}$  through a mapping process known as  $n$ -mode flattening, matricization or unfolding. To solve the BSS problem associated with the blind extraction of analytes, the mixtures data (1) will often have to be transformed into a new representation domain by means of some linear transform  $T$ :

$$T(\mathbf{X}) = \mathbf{A}T(\mathbf{S}) \quad (2)$$

Examples of such linear transform are wavelet or Fourier transforms.<sup>18</sup> The transform  $T$  is applied to  $\mathbf{X}$  row-wise. If mixtures are recorded by higher-dimensional spectroscopic or spectrometric modality (2D NMR for example) a higher-dimensional transform  $T$  is applied to each mixture before it is mapped to its one-dimensional counterpart. For example, the blind extraction of four analytes from three mixtures of 2D NMR spectra presented in the Results section has been carried out by transforming each mixture to a 2D wavelet domain to identify the matrix of concentrations, and then to a 2D Fourier domain to identify the spectrum of the analytes. The BSS concept for extraction of analytes requires a complex signal format in order to detect samples of single-component activity. This format is not supported for some modalities, such as FT-IR spectroscopy or mass spectrometry (MS). In such cases we propose a complex equivalent of real data  $\mathbf{X}$  that is obtained through the analytic representation:<sup>22</sup>

$$\tilde{\mathbf{X}} = \mathbf{X} + jH(\mathbf{X}) \quad (3)$$

where  $j = \sqrt{-1}$  denotes an imaginary unit and  $H$  denotes the Hilbert transform that is applied to  $\mathbf{X}$  row-wise. Since a complex format, such as (3), is only necessary to detect points (indicies) of the single analyte activity in the chosen basis, any transform that yields a complex signal format can be used for this purpose as well. We have used the analytic representation (3) in the experiment, reported in the Results section, related to the blind extraction of five analytes from two mixtures of MS data.

1  
2  
3 **Sparse representations and single-component-points.** The matrix factorization  $\mathbf{X}=\mathbf{A}\mathbf{S}$  assumed by  
4  
5 the linear mixture model (1) suffers from indeterminacies because  $\mathbf{A}\mathbf{T}\mathbf{T}^{-1}\mathbf{S}=\mathbf{X}$  for any invertible  $\mathbf{T}$ , i.e. it  
6  
7 implies that infinitely many  $(\mathbf{A},\mathbf{S})$  pairs can give rise to  $\mathbf{X}$ . The meaningful solution of the factorization  
8  
9 of  $\mathbf{X}$  is characterized with  $\mathbf{T}=\mathbf{P}\mathbf{\Lambda}$  where  $\mathbf{P}$  is the permutation matrix and  $\mathbf{\Lambda}$  is a diagonal matrix. These  
10  
11 standard blind decomposition indeterminacies are obtained by imposing statistical independence  
12  
13 constraints on  $\mathbf{S}$  when ICA<sup>12,23</sup> is used to solve related BSS problems.<sup>7-11</sup> As discussed previously, the  
14  
15 ICA-related requirements for analytes are not met when mixtures represent complex systems. They can  
16  
17 contain many analytes and it is therefore very likely that  $J>I_n$ . Therefore, a sparseness-based solution of  
18  
19 the BSS problem (1) is proposed. It is said that the  $n$ -dimensional signal  $\mathbf{y}$  is  $k$ -sparse in basis  $T$  if it is  
20  
21 represented by  $k\ll n$  coefficients, i.e. it is of special interest to look for the basis  $T$  where only a few  
22  
23 entries of the vector of coefficients  $T(\mathbf{y})$  are nonzero. In relation to the BSS problem associated with  
24  
25 model (1), we comment that the sparseness request applies to the  $J$ -dimensional column vectors  $\mathbf{s}_i$  of  $\mathbf{S}$   
26  
27 or  $T(\mathbf{S})$ ,  $i \in \{1, \dots, I_1 I_2 \dots I_{n-1}\}$ , while rows of  $\mathbf{S}$  or  $T(\mathbf{S})$  correspond to the analytes or their transformations.  
28  
29 However, it is clear that if the row vectors of  $\mathbf{S}$  or  $T(\mathbf{S})$  are sparse, the column vectors in the  
30  
31 corresponding representation will be sparse as well. In the absence of noise, if the column vectors of  
32  
33 either  $\mathbf{S}$  or  $T(\mathbf{S})$  are  $k=I_n-1$  sparse, i.e. have  $J-I_n+1$  zero components, a unique solution of the  
34  
35 underdetermined BSS (uBSS) problem, characterized with  $J>I_n$ , can be obtained.<sup>15</sup> Provided that the  
36  
37 column vectors of either  $\mathbf{S}$  or  $T(\mathbf{S})$  are  $k=1$  sparse, a unique solution of the uBSS problem can be  
38  
39 obtained, even from  $I_n=2$  mixtures only. However, for some signals such as those arising in NMR or FT-  
40  
41 IR spectroscopy it is very hard or even impossible to find a basis  $T$  where complex samples will be  $k=1$   
42  
43 sparse. Thus, instead of looking for the basis  $T$  that will yield  $k=1$  sparse representation of analytes at *all*  
44  
45 sample points (in the most general case there are  $I_1 I_2 \dots I_{n-1}$  sample points), we are interested in a  
46  
47 representation  $T$  that will provide us with only  $P$  sample points where analytes are  $k=1$  sparse such that  
48  
49  $J\leq P\ll I_1 I_2 \dots I_{n-1}$ . Since  $J\ll I_1 I_2 \dots I_{n-1}$  it ought to be possible to find such a small amount of points even  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

1 when analytes exhibit high degree of mutual similarity/complexity or their number is large. This belief is  
2 based on two facts: (i) the existence of a basis such as the wavelet basis with multiple degrees of  
3 freedom that provides signal representation at various resolution levels and different types of wavelet  
4 function; (ii) the number of detected points of single analyte activity is also governed by the choice of  
5 angular threshold  $\Delta\theta$  in the direction-based criterion (4), defined below. Thus, for the situation when  
6 analytes are highly complex or their number is large, the threshold  $\Delta\theta$  can be increased, slightly  
7 compromising accuracy of the method through detection not of components of single analyte activity but  
8 single analyte dominance. Yet, if the complexity of the analytes is very high or their number is large it  
9 might be necessary, for example, to use higher than 2D NMR spectroscopy. The use of points of single-  
10 component activity in the BSS has been exploited in DUET algorithm in ref. 24 for the separation of  
11 speech signals, wherein it has been assumed that at each point in the time-frequency plane, only one  
12 speech signal is active. In our approach, we rely on the geometric concept of direction to detect points  
13 where single analytes are present. This detection criterion was proposed in ref. 25. It requires complex  
14 representation of signals and was originally applied in the Fourier basis. The criterion is based on the  
15 notion that the real and imaginary parts of the complex vector of mixtures point either in the same or in  
16 opposite directions at the sample points of single analyte activity. This is based on the following  
17 reasoning. Let us denote by  $\mathbf{x}_i$  the complex column vector of either the mixtures data  $\mathbf{X}$  (1) or the  
18 transformed mixtures data  $T(\mathbf{X})$  (2) at the sample index  $i$ . At the point  $i$ , where only one analyte is  
19 active, it applies for the vector of mixtures:  $\mathbf{x}_i = \mathbf{a}_j s_{ij}$ , where  $\mathbf{a}_j$  is the vector of concentrations of the  $j$ th  
20 analyte across the mixtures and  $s_{ij}$  is the  $j$ th analyte that is active at point  $i$ . Since the vector of  
21 concentrations  $\mathbf{a}_j$  is real, the real and imaginary parts of vector  $\mathbf{x}_i$  must point in the same direction when  
22 the real and imaginary parts of  $s_{ij}$  have the same sign or, in opposite directions, when the real and  
23 imaginary parts of  $s_{ij}$  have different signs. Thus, the sample point  $i$  belongs to the set of single analyte  
24 points (SAPs) provided that the following criterion is satisfied  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60



$$\left| \frac{R\{\mathbf{x}_i\}^T I\{\mathbf{x}_i\}}{\|R\{\mathbf{x}_i\}\| \|I\{\mathbf{x}_i\}\|} \right| \geq \cos(\Delta\theta) \quad i \in \{1, \dots, I_1 I_2 I_{n-1}\} \quad (4)$$

where  $R\{\mathbf{x}_i\}$  and  $I\{\mathbf{x}_i\}$  denote the real and imaginary part of  $\mathbf{x}_i$ , respectively. 'T' denotes the transpose operation,  $\|R\{\mathbf{x}_i\}\|$  and  $\|I\{\mathbf{x}_i\}\|$  denote the  $\ell_2$ -norms of  $R\{\mathbf{x}_i\}$  and  $I\{\mathbf{x}_i\}$  and  $\Delta\theta$  denotes the angular displacement from directions of either 0 or  $\pi$  radians. Equation (4) follows from the definition of the inner product  $R\{\mathbf{x}_i\}^T I\{\mathbf{x}_i\} = \|R\{\mathbf{x}_i\}\| \|I\{\mathbf{x}_i\}\| \cos(\Delta\theta)$ . At single analyte points  $\Delta\theta=0$  and the inequality sign in (4) is replaced by an equality sign. Evidently, the smaller  $\Delta\theta$  is the smaller will be the number of candidates for identified as SAPs. However, the accuracy of the estimation of the number of analytes  $J$  and the concentration matrix  $\mathbf{A}$  will be greater. In this regard, when mixtures  $\mathbf{X}$  in (1) represent NMR signals we propose the use of the wavelet rather than the Fourier basis, to detect SAPs. If either the complexity of the analytes or their number is too great so that the chance of detecting sufficient SAPs is reduced (or zero),  $\Delta\theta$  can be increased. This, in part, will affect the accuracy of the estimation of the concentration matrix  $\mathbf{A}$  due to the fact that chance is increased that, instead of SAPs, we are detecting points where some of the analytes are dominant. This is important for not losing information about analytes that appear only on the diagonal in 2D NMR spectra and, therefore, are more likely to be dominant at a certain number of points rather than single one.

**Data clustering-based estimation of the number of analytes and the matrix of concentration profiles.** A satisfactorily identified set of the SAPs enables the accurate estimation of the number of analytes  $J$  and the matrix of concentration profiles  $\mathbf{A}$ . This is due to the fact that analytes in this set are  $k=1$  sparse and this condition, in the absence of noise, guarantees that the estimation of  $\mathbf{A}$  is unique up to the permutation and scale.<sup>15,25,26</sup> At SAPs the following relation holds:

$$\mathbf{x}_i = \mathbf{a}_j s_{j,i} \quad j \in \{1, \dots, J\}, \quad i \in \{1, \dots, I_1 I_2 \dots I_{n-1}\} \quad (5)$$

i.e. samples in the mixtures, which are column vectors of data matrix  $\mathbf{X}$ , coincide with some of the columns of  $\mathbf{A}$ . Thus,  $\mathbf{A}$  can be estimated from  $\mathbf{X}$  employing some of the available data-clustering algorithms.<sup>26,27</sup> However, in many BSS algorithms it is assumed that the number of analytes  $J$  is either known or can be estimated easily. This does not seem to be true in practice at all, especially when the BSS problem is underdetermined and mixtures represent samples of biological fluids or tissue extracts.<sup>3</sup> Generally speaking, the estimation of the number of analytes is a complex issue known in computer science as the intrinsic dimensionality problem.<sup>28</sup> A few related methods are described in references 29 to 31. However, they all assume  $J \geq I_n$ . Thus, they are not applicable to the uBSS problem that is of central interest here. To estimate the number of analytes for an identified set of SAPs, we propose to use the clustering function:<sup>16-18</sup>

$$f(\mathbf{a}) = \sum_{i=1}^P \exp\left(-\frac{d^2(\mathbf{x}_i, \mathbf{a})}{2\sigma^2}\right) \quad (6)$$

where  $d$  denotes the distance calculated as  $d(\mathbf{x}_i, \mathbf{a}) = \sqrt{1 - (\mathbf{x}_i \cdot \mathbf{a})^2}$  and  $(\mathbf{x}_i \cdot \mathbf{a})$  denotes the inner or dot product.  $\mathbf{a}$  represents the mixing vector in a two-dimensional subspace that is parameterized as:

$$\mathbf{a} = [\cos(\varphi) \quad \sin(\varphi)]^T \quad (7)$$

where  $\varphi$  represents the mixing angle that is confined in the interval  $[0, \pi/2]$  due to the non-negativity of the mixing coefficients (they represent concentration profiles of the analytes). Parameter  $\sigma$  defines the resolving power of the function  $f(\mathbf{a})$ . When  $\sigma$  is set to a sufficiently small value (in reported experiments this turned out to be  $\approx 0.05$ ) the value of the function  $f(\mathbf{a})$  will approximately equal the number of data

1 points close to  $\mathbf{a}$ . The number of peaks of the function  $f(\mathbf{a})$  in the interval  $[0, \pi/2]$  corresponds to the  
2  
3 estimate of the number of analytes  $J$  present in the mixtures. The selection of a two-dimensional  
4  
5 subspace out of an  $I_n$ -dimensional mixture space greatly simplifies the computational complexity of the  
6  
7 estimation process due to the fact that an  $(I_n-1)$  dimensional search in the space of mixing angles is  
8  
9 reduced to a one-dimensional search. The reduction to the two-dimensional subspace is enabled by the  
10  
11 fact that each analyte is present in some concentration in each of the  $I_n$  mixtures available. It is clear that  
12  
13 the value of  $\sigma$  reported above is empirical. For another set of mixtures it can yield a different value for  $J$ .  
14  
15 To obtain a robust estimator of the number of analytes  $J$ , we have proposed in ref. 16 to 18 to decrease  
16  
17 value of  $\sigma$  until the estimated number of analytes is increased by 1 or 2. False analytes will be either a  
18  
19 repeated version of some of the true analytes or their linear combinations. Thus, they can be detected  
20  
21 after blind extraction as the ones that are highly correlated with the rest of the extracted analytes. It is  
22  
23 also clear that if the concentration profiles of the analytes are very similar it will be increasingly more  
24  
25 difficult to discriminate them. In such a case, the solution might be to evaluate clustering function in 3D  
26  
27 or even higher-dimensional space, because this will decrease probability that different analytes have the  
28  
29 same concentration profiles across an increased number of mixtures. This however adds to the  
30  
31 computational complexity of the algorithm due to the fact that the one-dimensional search in the domain  
32  
33 of mixing angles is replaced by a search in a higher-dimensional space. After the number of analytes  $J$  is  
34  
35 estimated, the matrix of concentration profiles  $\mathbf{A}$  is estimated on the same set of SAPs employing some  
36  
37 of data clustering methods.<sup>27</sup> In the subsequent experimental setup, hierarchical and k-means clustering,  
38  
39 implemented through the *clusterdata* and *kmeans* commands from MATLAB's Statistical toolbox, have  
40  
41 been used for this purpose.  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51

52 **Estimation of analytes in over-, even- and under-determined scenarios.** When the estimated  
53  
54 number of analytes  $J$  is less than or equal to the number of mixtures  $I_n$ , the resulting BSS problem is  
55  
56 respectively over- or even-determined. Analytes can be estimated through the simple matrix pseudo-  
57  
58 inverse:  
59  
60

$$\mathbf{S}=\mathbf{A}^{\dagger}\mathbf{X} \quad (8a)$$

or

$$T(\mathbf{S})=\mathbf{A}^{\dagger}T(\mathbf{X}) \quad (8b)$$

where  $\mathbf{A}^{\dagger}$  denotes the Moore-Penrose pseudo-inverse of  $\mathbf{A}$ . Whether (8a) or (8b) is employed depends on the type of the spectroscopic modality that is used. If NMR spectroscopy is used it is customary to estimate analytes in the Fourier domain, in which case (8b) is preferred with  $T$  representing the Fourier transform (note that for NMR data  $\mathbf{A}$  is identified in the wavelet domain). If mass spectrometry or FT-IR spectroscopy are used it is customary to estimate analytes in the recording domain (8a). For reasons of clarity, we emphasize again that the transform  $T$  is applied to  $\mathbf{X}$  row-wise. If mixtures are recorded by higher-dimensional spectroscopic or spectrometric modality (2D NMR for example) a higher-dimensional transform  $T$  is applied to each mixture before it is mapped to its one-dimensional counterpart. The accuracy of the pseudo-inverse approach (8a)/(8b) with  $\mathbf{A}$  identified on a set of SAPs greatly outperforms the one obtained by ICA, as reported in ref. 25. When the number of analytes  $J$  is greater than the number of mixtures  $I_n$ , the resulting BSS problem is underdetermined. In such a case, the inverse problem has many solutions and simple pseudo-inverse approach (8a)/(8b) can no longer be applied. Provided that either  $\mathbf{s}_i$  or  $T(\mathbf{s}_i)$  are  $k=I_n-1$  sparse, i.e. have  $J-I_n+1$  zero components, it is possible to obtain the solution of the resulting uBSS problem through  $\ell_1$  norm minimization,<sup>14,15</sup> once the number of analytes  $J$  and concentration matrix  $\mathbf{A}$  are estimated. The analyte extraction problem is then reduced to solving the resulting underdetermined system of linear equations, that is carried out as linear programming<sup>14,32,33</sup> or the  $\ell_1$ -regularized least square problem.<sup>34,35</sup> Provided that the concentration matrix  $\mathbf{A}$  is estimated accurately, the result in ref. 23 states that the minimum of the  $\ell_1$  norm yields an accurate solution of the uBSS problem even if analytes are  $I_n$ -sparse, i.e. have  $J-I_n$  zero components. It means that  $I_n$  analytes can co-exist at each sample point. When multiple analytes occupy each sample

point of generally complex mixture, we respectively notice the relation between the real and imaginary parts of  $\mathbf{x}_i$  as:  $R\{\mathbf{x}_i\} = \mathbf{A}R\{\mathbf{s}_i\}$  and  $I\{\mathbf{x}_i\} = \mathbf{A}I\{\mathbf{s}_i\}$ ,  $i \in \{1, \dots, I_1 I_2 I_{n-1}\}$ . Written in matrix formulation it reads as

$$\begin{bmatrix} R\{\mathbf{x}_i\} \\ I\{\mathbf{x}_i\} \end{bmatrix} = \begin{bmatrix} \mathbf{A} & \mathbf{0} \\ \mathbf{0} & \mathbf{A} \end{bmatrix} \begin{bmatrix} R\{\mathbf{s}_i\} \\ I\{\mathbf{s}_i\} \end{bmatrix} \quad (9a)$$

or

$$\bar{\mathbf{x}}_i = \bar{\mathbf{A}} \bar{\mathbf{s}}_i \quad (9b)$$

where in (9a)  $\mathbf{0}$  is the matrix with the same dimensions as  $\mathbf{A}$  and all entries equal to 0. We introduce dummy variables  $\mathbf{u}, \mathbf{v} \geq \mathbf{0}$  such that  $\bar{\mathbf{s}}_i = \mathbf{u} - \mathbf{v}$ . Assuming that  $\mathbf{z} = \begin{pmatrix} \mathbf{u} \\ \mathbf{v} \end{pmatrix}$  and  $\bar{\bar{\mathbf{A}}} = [\bar{\mathbf{A}} \quad -\bar{\mathbf{A}}]$  linear programming based solution with equality constrains is obtained as

$$\hat{\mathbf{z}}_i = \arg \min_{\mathbf{z}_i} \sum_{j=1}^{2J} z_{j,i} \quad \text{subject to } \bar{\bar{\mathbf{A}}} \mathbf{z}_i = \bar{\mathbf{x}}_i \quad \forall i = 1, \dots, I_1 I_2 I_{n-1} \\ \mathbf{z}_i \geq \mathbf{0} \quad (10)$$

Linear programming (10) favors solution with the minimal  $\ell_1$  norm. With high probability, this is the sparsest solution of (9b).<sup>33-35</sup> Hence, if analytes satisfy the desired degree of mutual sparseness, the  $k \leq I_n$  solution of (10) will successfully recover them. Analytes are obtained from the solution of the linear program (10) as  $\bar{\mathbf{s}}_i$ , where  $\mathbf{u}$  is obtained from the upper half of  $\hat{\mathbf{z}}_i$  and  $\mathbf{v}$  is obtained from the lower half of  $\hat{\mathbf{z}}_i$ . The real part of  $\mathbf{s}_i$  is obtained from the upper half of  $\bar{\mathbf{s}}_i$  while the imaginary part of  $\mathbf{s}_i$  is obtained from the lower half of  $\bar{\mathbf{s}}_i$ . If noise is present in the uBSS problem, a more robust solution for  $\hat{\mathbf{z}}_i$  (thus also  $\mathbf{s}_i$ ) is obtained by solving the  $\ell_1$ -regularized least square problem.<sup>34</sup>

$$\hat{\mathbf{z}}_i = \arg \min_{\mathbf{z}_i} \frac{1}{2} \left\| \overline{\mathbf{A}} \mathbf{z}_i - \overline{\mathbf{x}}_i \right\|_2^2 + \lambda \left\| \mathbf{z}_i \right\|_1 \quad \forall i = 1, \dots, I_1 I_2 I_{n-1} \quad (11)$$

Solution of (11) minimizes the  $\ell_2$ -norm of the error between data  $\overline{\mathbf{x}}_i$  and its model  $\overline{\mathbf{A}} \mathbf{z}_i$ , trading the degree of error for the degree of sparseness of the solution. The degree of compromise is balanced by the value of the regularization factor  $\lambda$ . There are other methods developed over last few years for solving underdetermined systems of linear equations. Most notable are methods that minimize the  $\ell_p$ -norm ( $0 < p \leq 1$ ) of the solution coefficients (analytes) such as the iterative recursive least square (IRLS) algorithm,<sup>36</sup> methods that optimize null-space of the concentration matrix  $\mathbf{A}$ ,<sup>37</sup> and methods that work with a smooth approximation of the  $\ell_0$ -quasi norm of the solution coefficients.<sup>38</sup> We have checked the IRLS algorithm and the smoothed  $\ell_0$ -quasi norm algorithm on the experimental problem considered below. These methods did not bring any improvement relative to the performance achieved by the interior point method employed to solve the  $\ell_1$ -regularized least square problem (11) or the linear programming method employed to solve (10).

## EXPERIMENTAL SECTION

**NMR measurements.** We used 6-*O*-(*N,O*-bis-*tert*-butyloxycarbonyl-L-tyrosyl-L-prolyl)-D-glucopyranose (**1**), 6-*O*-(*N,O*-bis-*tert*-butyloxycarbonyl-L-tyrosyl-L-prolyl-L-phenylalanyl)-D-glucopyranose (**2**), 6-*O*-(*N-tert*-butyloxycarbonyl-L-prolyl-L-phenylalanyl-L-valyl)-D-glucopyranose (**3**) and 6-*O*-(*N,O*-bis-*tert*-butyloxycarbonyl-L-tyrosyl-L-prolyl-L-phenylalanyl-L-valyl)-D-glucopyranose (**4**)<sup>39</sup> to prepare three mixtures with different ratios of **1-4**:  $\mathbf{X}_1$  (**1:2:3:4** = 1.1:1.7:2.7:1),  $\mathbf{X}_2$  (**1:2:3:4** = 2.5:1.7:1.3:1) and  $\mathbf{X}_3$  (**1:2:3:4** = 1:4:2.7:2.2). To test the ability of the ICA-based approach, which requires the number of mixtures to be equal or greater than the number of analytes, the fourth mixture  $\mathbf{X}_4$  (**1:2:3:4** = 3.2:1:2.3:3.5) has been prepared and treated as described above. Compounds **1-4** and mixtures  $\mathbf{X}_1$ - $\mathbf{X}_4$  were dissolved in 600  $\mu\text{L}$  of DMSO- $d_6$  and NMR spectra recorded with a Bruker

1 AV300 spectrometer, operating at 300.13 MHz and 298 K. The  $^1\text{H}$ - $^1\text{H}$  COSY spectra were obtained in  
2 the magnitude mode with 2048 points in the  $F_2$  dimension and 512 increments in the  $F_2$  dimension.  
3  
4 Each increment was obtained with 4 scans and a spectral width of 6173 Hz. The resolution was 3.01 and  
5  
6 6.02 Hz per point in the  $F_1$  and  $F_2$  dimensions, respectively.  
7  
8

9  
10  
11 **Mass spectrometry measurements.** The compounds used for the analysis and procedures regarding  
12 MS measurements are described in ref. 17.  
13  
14

15  
16  
17 **Software environment.** The BSS method described was tested on the decomposition of 2D COSY  
18 NMR spectra and mass spectra using custom scripts in the MATLAB programming language (version  
19 7.1.; The MathWorks, Natick, MA). The data clustering part of the SCA algorithm was implemented  
20 using the *clusterdata* and *kmeans* commands from the Statistics toolbox. The *clusterdata* command was  
21 used with the following set of parameters: distance: cosine; linkage: complete; maxclus: J, where J  
22 represents number of analytes estimated previously from the peaks of the clustering function (6)/(7). The  
23 linear programming part of the SCA algorithm was implemented using the *linprog* command from the  
24 Optimization toolbox and the interior point method.<sup>34,40</sup> The two-dimensional wavelet transform was  
25 implemented using the *swt2* command from the Wavelet toolbox. All programs were executed on a PC  
26 running under the Windows XP operating system using an Intel Core 2 Quad Processor Q6600 operating  
27 with a clock speed of 2.4 GHz and 4GB of RAM installed.  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46

## 47 RESULTS AND DISCUSSION

### 48 **Setting up an experiment**

49 To demonstrate the efficiency of the proposed multivariate data analysis method, a "control" experiment  
50 was set up. Since we are targeting *complex mixtures*, it was important to choose a group of compounds  
51 that will comply with complexity requirement. It was also important to have known and well-  
52 characterized compounds to verify accuracy of estimation. Among many options, we have selected the  
53  
54  
55  
56  
57  
58  
59  
60

glycopeptides **1-4**,<sup>39</sup> where the N-terminally protected dipeptide (Tyr-Pro, **1**), tripeptides (Tyr-Pro-Phe, **2**) and (Pro-Phe-Val, **3**) and tetrapeptide (Tyr-Pro-Phe-Val, **4**) are linked to the C-6 group of D-glucose (Figure 1a). Crude compounds **1-4** were mixed to obtain three mixtures with different concentrations of components (see Experimental Section for details).

As seen from their structures, compounds **1-4** are structurally analogous, and consequently their spectral profiles are, to a large extent, similar (Figure 1b). Additionally, the presence of a reducing sugar gives rise to both  $\alpha$ - and  $\beta$ -pyranose forms in the solution, while the presence of the proline residue causes *cis-trans* isomerization of the X-Pro peptide bond. All together, accurate assignment of all resonances requires 2D NMR measurements. Even so, COSY spectra obtained from mixtures consisted of compounds **1-4** (Figure 2) showed overlapping, which undoubtedly hampered assignment. Thus, the proposed mixture model passed the complexity requirement and seemed adequate for the data analysis.

**Blind extraction of four analytes from three mixtures in 2D NMR spectroscopy.** Figures 1 to 4 and Table 1 demonstrate the experimental blind extraction of four pure-component COSY spectra from three mixtures by means of the described sparseness-based multivariate data analysis method. The COSY spectra of compounds **1-4** are presented in Figure 1b, while Figure 2 shows the COSY spectra of the three mixtures. The structural similarity of the selected compounds accounts for the complexity and overlapping in the sugar resonance region, as well as in the amino-acid amide and side-chain resonance areas. A convenient way to quantify this overlapping is to calculate normalized correlation coefficients between the spectra of the analytes **1-4** (Table 1a). It is clear that compound **1** is highly correlated with (similar to) compound **2** (0.5509). Compound **2** is additionally correlated with **4** (0.5120), while compounds **3** and **4** are highly mutually correlated, with a coefficient of 0.7965. Clearly, these correlation coefficients reflect structural and spectral similarities between the studied compounds and allow simplified numerical analysis of often complex NMR spectra.

Clustering functions described by the equations (6) and (7), are shown in the mixing angle domain in Figure 3 for three two-dimensional subspaces  $\mathbf{X}_1\mathbf{X}_2$ ,  $\mathbf{X}_1\mathbf{X}_3$  and  $\mathbf{X}_2\mathbf{X}_3$ , i.e. all combinations of two



1 mixtures were used for the estimation of the number of analytes present in the mixtures. The clustering  
2 functions were calculated on a set of 203 SAPs, (equation (5)), detected in the symmlet 8 wavelet  
3 domain using direction based criterion (4) with angular displacement set to  $\Delta\theta=1$  degree. The value of  
4 the dispersion factor  $\sigma$  in (5) has respectively been set to 0.04, 0.06 and 0.05. The meaning of 203 SAPs  
5 is that only one of components **1-4** was active at only 203 out of 65536 points available. The four peaks  
6 in clustering functions suggest the existence of four analytes in the mixtures. The small variation of the  
7 dispersion factors parameter confirms the statement that any two-dimensional mixtures subspace can be  
8 used for the estimation of the number of analytes.  
9

10 The spectra of the pure components estimated from three mixtures  $\mathbf{X}_1$ - $\mathbf{X}_3$  (Figure 2) are shown in  
11 Figure 4. Since the concentration matrix is estimated accurately on a subset of SAPs, the  $\ell_1$ -regularized  
12 least square method, eq.(11), yielded good estimates of the analytes spectra, even when two components  
13 occupy the same frequency. The similarity between the spectra of pure and estimated analytes is  
14 quantified in Table 1b, where normalized correlation coefficients between the true and estimated  
15 analytes spectra are shown. The closer these numbers are to those in Table 1a, the better is the extraction  
16 of the components from the mixtures. Inspection of the data shows that all four components were  
17 successfully separated from three mixtures; even highly correlated components **3** and **4** are assigned  
18 reliably.  
19

20 To demonstrate the importance of the wavelet basis for providing sparse representation of the NMR  
21 signals, we have estimated the set of SAPs in the Fourier domain with angular displacement criterion set  
22 to  $\Delta\theta=2$  degrees, i.e. two times greater than in the case of wavelet basis. However, only 23 SAPs were  
23 detected in this case and the estimation of the matrix of concentration profiles was less accurate.  
24 Consequently, the  $\ell_1$ -regularized least square method failed to provide good estimates of the analytes  
25 spectra. This is quantified in Table 1c, where normalized correlation coefficients between pure analytes  
26 spectra and spectra of the estimated analytes in Fourier domain are shown. While the accuracy of the  
27 estimation of components **1-3** generally follows that in Table 1b, component **4** is estimated incorrectly.  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

1 This is a consequence of a high degree of similarity (correlation factor) in combination with a low  
2 number of SAPs detected in the Fourier domain. Therefore, the importance of the wavelet basis for  
3 providing sparse representation of the NMR signals is clearly verified.  
4  
5

6  
7 Finally, significant degrees of correlation between spectra of the pure analytes would cause the  
8 ICA-based approach to fail even if the number of mixtures would be equal to the number of analytes.  
9 This is due to the fact that significant correlation between spectra of the pure analytes violates the  
10 statistical independence assumption required by ICA. This has been demonstrated by using the JADE  
11 ICA algorithm<sup>41</sup> to separate the same four analytes but from four mixtures, whereupon normalized  
12 correlation coefficients between pure and estimated analytes spectra are shown in Table 1d.  
13  
14  
15  
16  
17  
18  
19  
20

21 As discussed previously, only a few methods or algorithms have been developed for the extraction of  
22 analytes from multicomponent spectral data without any known *a priori* information. The majority of  
23 these blind decomposition methods require the number of mixtures to be greater than or equal to the, in  
24 principle *unknown*, number of pure components. Some of these methods, like the band-target entropy  
25 minimization (BTEM), have been applied on the extraction of components in 2D NMR (COSY and  
26 HSQC) spectroscopy from multicomponent mixtures.<sup>13</sup> However, *seven* mixtures were used for the  
27 reconstruction of *three* pure components of simple structure. Moreover, as stated by the authors, the  
28 BTEM approach is inapplicable when the number of experimentally measured spectra is less than the  
29 number of observed components. Here, as well as in recent publications,<sup>16-18</sup> we have demonstrated that  
30 the sparseness-based approach successfully estimates pure components when the number of available  
31 mixtures is less than the unknown number of components.  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48

49 **Blind extraction of five analytes from two mixtures in mass spectrometry.** The proposed  
50 sparseness-based multivariate data analysis method for blind analyte extraction relies on the detection of  
51 a large enough set of SAPs in a suitably chosen basis, using direction-based criterion (4). As discussed  
52 previously, the direction based criterion (4) requires complex signals. To circumvent this difficulty for  
53 the case of real signals, arising for example in mass spectrometry or FT-IR spectroscopy, we have  
54  
55  
56  
57  
58  
59  
60

1 proposed the use of the analytic representation (3) of the real signals to detect the positions of the SAPs.  
2  
3 We have recently described blind extraction of five pure components mass spectra from only two  
4  
5 mixtures by means of sparse component analysis.<sup>17</sup> The structures of the pure components, their mass  
6  
7 spectra and the mass spectra of the two mixtures are available as Supporting information (Figures S-1,  
8  
9 S-2 and S-3, respectively). The same data set was used to validate the sparseness-based multivariate data  
10  
11 analysis method proposed herein. By setting the angular displacement criterion to  $\Delta\theta = 2$  degrees, 290  
12  
13 SAPs were detected using the analytical representation (3). The clustering function (Figure S-4) showed  
14  
15 five peaks corresponding to five analytes present in the mixtures. The estimated mass spectra are  
16  
17 presented in Figure S-5 and are consistent with the results already obtained in Ref. 17. This is due to the  
18  
19 fact that the mass spectra of the analytes were weakly correlated (see Table S-1 in the supporting  
20  
21 information). However, this validates an approach for the detection of SAPs in the case of real signals,  
22  
23 which is based on the use of an analytic representation (3).  
24  
25  
26  
27

28  
29 This result should be considered in the wider context of the utility of mass spectrometry for metabolic  
30  
31 profiling. Chromatographic separation of analytes present in mixtures prior to MS analysis is a standard  
32  
33 procedure, but suffers from some drawbacks. Different samples (mixtures) require different separation  
34  
35 techniques (column packages, mobile phases) and determining optimal conditions for the separation is  
36  
37 usually a time- and resource-consuming process.<sup>3</sup> On the other hand, direct infusion of the complex  
38  
39 sample into the mass spectrometer is generally not applicable, owing to the ionization suppression and  
40  
41 the formation of adducts in the ion source.<sup>1</sup> There are a few successful examples that are, however,  
42  
43 limited to the analysis of plant extracts.<sup>1</sup> The presented multivariate data analysis method based on the  
44  
45 detection of SAPs can reduce the need for the accurate separation prior to MS analysis and represents an  
46  
47 innovative approach for the metabolic analysis based on mass spectrometry. Furthermore, we plan to test  
48  
49 this approach on less "controlled" and more biologically relevant experiments, to determine possibilities  
50  
51 and limitations of the presented multivariate data analysis method.  
52  
53  
54  
55  
56  
57  
58  
59  
60

## Conclusion

1  
2  
3 We developed and demonstrated a sparseness-based method for blind estimation of analytes exhibiting a  
4  
5 high level of complexity and structural similarity, whereupon their number is greater than the number of  
6  
7 mixtures available. The method relies on the realistic assumption about the existence of a representation  
8  
9 domain or basis where a small number of data sample points can be found at which analytes do not  
10  
11 overlap. Although of general importance, the method was developed to solve an important problem in  
12  
13 metabolic studies: the blind extraction of analytes from a possibly smaller number of mixtures of NMR  
14  
15 or mass spectra. We exemplified the method through the estimation of four analytes from three mixtures  
16  
17 in 2D NMR spectroscopy and five analytes from two mixtures in mass spectrometry. The advantages of  
18  
19 the proposed sparseness-based approach over the presently used multivariate data analysis methods are  
20  
21 expected to be of greatest significance in applications such as metabolic profiling of biological fluids  
22  
23 and tissues in search for new biomarkers, analysis of plant and microbial extracts in seeking new  
24  
25 biologically active compounds and the reconstruction of transcription factors in gene regulating  
26  
27 networks.  
28  
29  
30  
31  
32  
33  
34  
35  
36

37 **Acknowledgment.** The work of I. Kopriva and I. Jerić was respectively supported by the Ministry of  
38  
39 Science, Education and Sports, Republic of Croatia under Grants 098-0982903-2558 and 098-0982933-  
40  
41 2936. David Smith's help in proofreading the manuscript is also gratefully acknowledged.  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52

53 **Supporting information available:** Structures of analytes **5-9**, Figure S-1; mass spectra of analytes **5-**  
54  
55 **9**, Figure S-2; mass spectra of two mixtures, Figure S-3; clustering function calculated on a set of 290  
56  
57 SAPs using analytic representation (3), Figure S-4; mass spectra of five estimated analytes, Figure S-5;  
58  
59  
60

1 normalized correlation coefficients for pure analytes **5-9** and analytes **5-9** estimated on 290 SAPs  
2  
3 detected by using analytical representation (3), Table S-1.  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60

## FIGURE CAPTIONS

**Figure 1.** a) Structures of glycopeptides **1-4**; b) COSY NMR spectra of pure analytes **1-4**.

**Figure 2.** COSY NMR spectra of three mixtures  $\mathbf{X}_1\text{-}\mathbf{X}_3$ .

**Figure 3.** Clustering functions calculated on 203 SAPs in wavelet domain for three two-dimensional mixture subspaces:  $\mathbf{X}_1\mathbf{X}_2$ ,  $\mathbf{X}_1\mathbf{X}_3$  and  $\mathbf{X}_2\mathbf{X}_3$ . Positions of four peaks P1 to P4 in each function are marked.

**Figure 4.** COSY NMR spectra of estimated analytes **1-4**.

## TABLE CAPTION

**Table 1.** Normalized correlation coefficients for (a) pure analytes **1-4**; (b) analytes **1-4** estimated on 203 SAPs detected in symmlet 8 wavelet domain; (c) analytes **1-4** estimated on 23 SAPs detected in Fourier domain; (d) analytes **1-4** estimated by means of JADE ICA algorithm from four mixtures.\* Significant degree of correlation between spectra of true analytes caused failure of the ICA-based extraction of analytes, Table 1d.

**References**

- 1  
2  
3 (1) van der Greef, J.; Stroobant, P.; van der Heijden, R. *Curr. Opinion Chem. Biol.* **2004**, *8*, 559-565.
- 4  
5  
6 (2) Ellis, D. E.; Dunn, W. B.; Griffin, J. L.; Allwood, J. W.; Goodacre, R. *Pharmacogenomics* **2007**,  
7  
8 8, 1243-1266.
- 9  
10  
11 (3) Lindon, J. C.; Nicholson, J. K. *Ann. Rev. Anal. Chem.* **2008**, *1*, 45-69.
- 12  
13  
14 (4) Méric, P.; Autret, G.; Doan, B. T.; Gillet, B.; Sébrié, C.; Beloeil, J.-C. *Magn. Reson. Mat. Phys.*  
15  
16 *Biol. Med.* **2004**, *17*, 317-338.
- 17  
18  
19  
20 (5) Thomas, M. A.; Lange, T.; Velan, S. S.; Nagarajan, R.; Raman, S.; Gomez, A.; Margolis, D.;  
21  
22 Swart, S.; Raylman, R. R.; Schulte, R. F.; Boesiger, P. *Magn. Reson. Mat. Phys. Biol. Med.* **2008**, *21*,  
23  
24 443-458.
- 25  
26  
27  
28 (6) Ye, T.; Mo, H.; Shanaiah, N.; Gowda, G. A. N.; Zhang, S.; Raftery, D. *Anal. Chem.* **2009**, *81*,  
29  
30 4882-4888.
- 31  
32  
33 (7) Nuzillard, D.; Bourg, S.; Nuzillard, J. M. *J. Mag. Res.* **1998**, *133*, 358-363.
- 34  
35  
36 (8) Visser, E.; Lee, T. W. *Chemometrics Intell. Lab. Syst.* **2004**, *70*, 147-155.
- 37  
38  
39 (9) Ren, J. Y.; Chang, C. Q.; Fung, P. C. W.; Shen, J. G.; Chan, F. H. Y. *J. Mag. Res.* **2004**, *166*, 82-  
40  
41 91.
- 42  
43  
44 (10) Shashilov, V. A.; Xu, M.; Ermolenkov, V. V.; Lednev, I. K. *J. Quant. Spectrosc. Radiat. Transf.*  
45  
46 **2006**, *102*, 46-61.
- 47  
48  
49 (11) Shao, X.; Wang, G.; Wang, S.; Su, Q. *Anal. Chem.* **2004**, *76*, 5143-5148.
- 50  
51  
52  
53 (12) Cichocki, A.; Amari, S.I. *Adaptive Blind Signal and Image Processing*; John Wiley: New York,  
54  
55 2002.
- 56  
57  
58  
59  
60

- 1 (13) Guo, L.; Wiesmath, A.; Sprenger, P.; Garland, M. *Anal. Chem.* **2005**, *77*, 1655-1662.  
2  
3  
4 (14) Bofill, P.; Zibulevsky, M. *Sig. Proc.* **2001**, *81*, 2353-2362.  
5  
6  
7 (15) Georgiev, P.; Theis, F.; Cichocki, A. *IEEE Trans. Neural Net.* **2005** *16*, 992-996.  
8  
9  
10 (16) Kopriva, I.; Jerić, I.; Cichocki, A. *Chemometrics Intell. Lab. Syst.* **2009**, *97*, 170-178.  
11  
12  
13 (17) Kopriva, I.; Jerić, I. *J. Mass Spect.* **2009**, *44*, 1378-1388.  
14  
15  
16 (18) Kopriva, I.; Jerić, I.; Smrečki, V. *Anal. Chim. Acta.* **2009**, *653*, 143-153.  
17  
18  
19 (19) Kitano, H. *Science* 2002, *295*, 1662-1664.  
20  
21  
22 (20) Liao, J. C.; Boscolo, R.; Yang, Y.-L.; Tran, L. M.; Sabatti, C.; Roychowdhury, V. P. *PNAS* **2003**,  
23 *100*, 15522-15527.  
24  
25  
26 (21) Kiers, H. A. L. *J. Chemometr.* **2000**, *14*, 105-122.  
27  
28  
29 (22) Gabor, D. *Trans. Inst. Electr. Eng.* **1946**, *93*, 429-456.  
30  
31  
32 (23) Comon, P. *Sig. Proc.* **1994**, *36*, 287-314.  
33  
34  
35 (24) Jourjine, A.; Rickard, S.; Yilmaz, O. in *Proc. Int. Conf. on Acoust., Speech, Signal Processing*,  
36 **2000**, *5*, pp. 2985-2988.  
37  
38  
39 (25) Reju, V. G.; Koh, S. N.; Soon, I. Y. *Sig. Proc.* **2009**, *89*, 1762-1773.  
40  
41  
42 (26) Naini, F. M.; Mohimani, G. H.; Babaie-Zadeh, M.; Jutten, Ch. *Neurocomputing* **2008**, *71*, 2330-  
43 2343.  
44  
45  
46 (27) Gan, G.; Ma, Ch.; Wu, J. *Data Clustering - Theory, Algorithms and Applications*; SIAM:  
47 Philadelphia, 2007.  
48  
49  
50 (28) Fukunaga, K.; Olsen, D.R. *IEEE Trans. Comput.* **1971**, *C-20*, 176-183.  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60



- 1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60
- (29) Malinowski, E. R. *Anal. Chem.* **1977**, *49*, 612-617.
- (30) Levina, E.; Wagman, A. S.; Callender, A. F.; Mandair, G. S.; Morris, M. D. *J. Chemometr.* **2007**, *21*, 24-34.
- (31) Westad, F.; Kermit, M. *Anal. Chim. Acta.* **2003**, *490*, 341-354.
- (32) Takigawa, I.; Kudo, N.; Toyama, J. *IEEE Trans. Signal Proc.* **2004**, *52*, 582-591.
- (33) Donoho, D. L.; Elad, M. *Proc. Nat. Acad. Sci. USA*, **2003**, *100*, 2197-2202.
- (34) Kim, S. J.; Koh, K.; Lustig, M.; Boyd, S.; Gorinevsky, S. *IEEE J. Sel. Topics Signal Proc.* **2007**, *1*, 606-617.
- (35) Tropp, J. A.; Gilbert, A. C. *IEEE Trans. Inf. Theory.* **2007**, *53*, 4655-4666.
- (36) Cahrtrand, R.; Staneva, V. *Inverse Problems.* **2008**, *24*, 035020 (14pp).
- (37) Kim, S.G.; Yo, Ch.D. *IEEE Trans. Signal Proc.* **2009**, *57*, 2604-2614.
- (38) Mohimani, H.; Babaie-Zadeh, M.; Jutten, C. *IEEE Trans. Signal Proc.* **2009**, *57*, 289-301.
- (39) Jerić, I.; Horvat, Š. *Eur. J. Org. Chem.* **2001**, 1533-1539.
- (40) [http://www.stanford.edu/~boyd/l1\\_ls/](http://www.stanford.edu/~boyd/l1_ls/).
- (41) Cardoso, J. F.; Soulomiac, A. *Proc. IEE F.* **1993**, *140*, 362-370.







Figure 1



58  
59  
60

Figure 2




Figure 3




Figure 4

Table 1:

entry	$An_1$	$An_2$	$An_3$	$An_4$	
a	$An_1$	<b>1</b>	0.5509	0.1394	0.3730
	$An_2$	0.5509	<b>1</b>	0.3051	0.5120
	$An_3$	0.1394	0.3051	<b>1</b>	0.7965
	$An_4$	0.3730	0.5120	0.7965	<b>1</b>
b	$\hat{An}_1$	<b>0.8931</b>	0.4753	0.2638	0.4132
	$\hat{An}_2$	0.5634	<b>0.8579</b>	0.2795	0.5366
	$\hat{An}_3$	0.1945	0.5048	<b>0.8990</b>	0.7953
	$\hat{An}_4$	0.4386	0.6124	0.8060	<b>0.8381</b>
c	$\hat{An}_1$	<b>0.8924</b>	0.6009	0.2754	0.4602
	$\hat{An}_2$	0.5482	<b>0.8469</b>	0.3107	0.5695
	$\hat{An}_3$	0.0931	0.4101	<b>0.8432</b>	0.7249
	$\hat{An}_4$	0.3108	0.3411	<b>0.8236</b>	0.7331
d	$\hat{An}_1$	0.7189	0.7090	0.6805	<b>0.7939</b>
	$\hat{An}_2$	0.6873	0.7571	0.6524	<b>0.7790</b>
	$\hat{An}_3$	0.6606	0.7325	0.7142	<b>0.8177</b>
	$\hat{An}_4$	0.6322	0.7232	0.7474	<b>0.8342</b>

\*  $An_1$ - $An_4$  pure analytes **1-4**;  $\hat{An}_1$ - $\hat{An}_4$  estimated analytes **1-4**.